

Checking record



Establishment TECHNOSKLO s.r.o. Držkov 135 applied and using system of quality according to the DIN EN ISO 9 001: 2015, certificate made out by TÜV Rhainland FRG


Sort of product: **DNR005 (632 436 104 619) – AUTOMATIC BURETTE ACCORDING TO SCHILLING, WITH ACCESSORIES, BOTTLE 500 ML**

Nominal volume: **10 ml**
Division: **0,05 ml**
Permitted volume error: **± 0,050 ml**

Testing instrument: balance – KERN ABT 220-4M; id. number: 34 / 07
 thermometer – enclosed – scale, (0 – 50)°C 0,01°C; id. number 1 / 02
Used standards are in accordance to standards of State Metrology.

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters according to the drawing No: SV 4275

Date of edition: 03.11.2020

Confirmed by: 
manager of quality