



CERTIFICATE OF CONFORMITY

Issuer's Name / Producer : GLASSCO LABORATORY EQUIPMENTS PVT. LTD.

Issuer's Address / Producer : Manglai, P.O. Khudda Kalan, Ambala Cantt - 133001 Haryana, India

Object Of Declaration : All-Glass Filter Holder, for 47mm Disc Filter

Material : Borosilicate Glass

Purpose of Use : Glassco filtration kits are used in vacuum filtration of liquids for analysis of particulate or biological contamination retained on the filter surface.

Specification of Filtration Kit as Glassco Specification:

Catalogue No.	Product Description
258.202.01*	All-Glass Filter Holder Assembly With Funnel, Fritted Base, Cap, Clamp, 47mm
Replacement Parts:	
258.245.01*	Glass Funnel, 300mL., Borosilicate
258.245.02*	Spring Clamp, 47mm, Aluminium
258.245.03*	Glass Base & Cap, 47mm
258.245.04*	Ground Joint Flask, 1000mL.



The object of declaration described above is in conformity with the requirements of the following documents:

- DIN ISO 3585 Borosilicate Glass 3.3
- * Indicates non DIN & non ASTM standards.

Material: The above are made from DIN ISO 3585:1998, heat resistant Boro 3.3 glass that meet ASTM E-438, TYPE-1, CLASS- A standard.

Compliance: These Filtration Kit comply in every respect characteristics, measures and accuracy are in full Conformity With the provisions of Glassco Specification

Flow Rate: ROUND FRIT 40-60 microns COURSE MEDIA (FLOW RATE 5-15 SEC/500MLS @ 27 INCH OF MERCURY REF.)
ROUND FRIT 40-60 microns COURSE MEDIA (FLOW RATE 5-30 SEC/500MLS @ 27 INCH OF MERCURY REF.)

Additional Information:

Glassco has installed the quality control system as per ISO 9001 and products delivered to the market are in full conformity with the technical Specification as per standard and with all fundamental requirements to such products.

Certificate No. IPK 0039.69-01 issued by intertek certification limited, A UKAS Management System Accredited Body



This document has been produced electronically & is valid without signature