



# CERTIFICATE OF CONFORMITY

Issuer's Name / Producer : GLASSCO LABORATORY EQUIPMENTS PVT. LTD.  
Issuer's Address / Producer : Manglai, P.O. Khudda Kalan, Ambala Cantt - 133004  
Haryana, India  
Object of Declaration : Columns  
Material : Borosilicate Glass  
Purpose of Use : The Column is used in rectification, also commonly called fractional  
Distillation, for improved thermal separation of liquid mixture.

## Specification of Vigreux Columns specified in Table-1.

Catalogue No.	Product Description	Socket Size	Cone Size	Effective Length (mm)	Pack Qty.
197.204.01	Column acc. To Vigreux Effective Length	10/19	10/19	100	1
197.204.01A	Column acc. To Vigreux Effective Length	14/23	14/23	100	1
197.204.02	Column acc. To Vigreux Effective Length	10/19	10/19	200	1
197.204.03	Column acc. To Vigreux Effective Length	29/32	29/32	360	1
197.204.03A	Column acc. To Vigreux Effective Length	24/29	24/29	360	1
197.204.03B	Column acc. To Vigreux Effective Length	19/26	19/26	300	1
197.204.04	Column acc. To Vigreux Effective Length	29/32	29/32	400	1
197.204.05A	Column acc. To Vigreux Effective Length	14/23	14/23	300	1
197.204.05C	Column acc. To Vigreux Effective Length	24/29	24/29	300	1
197.204.05D	Column acc. To Vigreux Effective Length	29/32	29/32	300	1
197.204.05E	Column acc. To Vigreux Effective Length	29/32	29/32	320	1
197.204.06	Column acc. To Vigreux Effective Length	24/29	24/29	600	1
197.204.06A	Column acc. To Vigreux Effective Length	29/32	29/32	600	1

**The object of declaration described above is in conformity with the requirements of the following documents:**

- DIN ISO 3585** **Borosilicate Glass 3.3**

**Material:** The above are made from DIN ISO 3585:1998, heat resistant Boro 3.3 Glass that meet ASTM E-438, TYPE-1, CLASS- A Standard.

### **Additional Information:**

Glassco has installed the quality control system as per ISO 9001 and products delivered to the market are in full conformity with the technical Specification as per standard and with all fundamental requirements to such products.

**Certificate No. IPK 0039.06 issued by intertek certification limited, A UKAS Management System Accredited Body**

This document has been produced electronically & is valid without signature

